

CERTIFICATION

STERILITY

Sterets are not terminally sterilised and are not marketed as Sterile. However, the solution is passed through a 0,2 micron filter before impregnating the viscose pad/swab.

PYROGENS

The Steret swab is intended for topical use on intact skin, so not tested for pyrogens.

We routinely test the Purified Water used in its manufacture.

The specification is Not More Than 0,025 EU/ml and we have not failed to meet this to date.

TOXICITY

The solution contains only Purified Water and Isopropyl Alcohol, both to EP standards.

The toxicity of IPA is well documented. The solution contains 70% v/v IPA which has well known antiseptic properties.

QUALITY

Sterets are manufactured in accordance with the CE requirements

REF no.	Lot.No.	Quantity	Exp. date